



Endosafe[®]
LAL Grade Glassware (TL700)
Certificate of Quality

Charles River Laboratories, Inc. 16 x 90 mm Borosilicate Glass Test Tubes with Screw Caps, 70/Pack Lot Number 94311C has been depyrogenated in a dry heat oven using a validated process. The lot has been tested by Quality Control and meets the following test specifications:

Endotoxin Test Result: <0.005 EU/device

Interference Test Result: Pass

Expiration Date: Jan 2022

Checked By Paul Angert Date 01 Feb 2019

Reviewed By Donna Sigl Date 05 Feb 2019

Charles River Laboratories, Inc.
1023 Wappoo Road, Suite 43-B
Charleston, SC 29407 USA