



## CERTIFICATE OF ANALYSIS

|                             |                          |                          |                |
|-----------------------------|--------------------------|--------------------------|----------------|
| <b>Catalog Number:</b>      | ECV2500V                 | <b>Lot Number:</b>       | 26055158       |
| <b>Product:</b>             | Endotoxin Challenge Vial | <b>Manufacture Date:</b> | 17-May-2026    |
| <b>Bacterial strain</b>     | <i>E.coli</i> O111: B4   | <b>Expiration Date:</b>  | 16-May-2028    |
| <b>Contents</b>             | <b>Description</b>       | <b>Catalog No.</b>       | <b>Lot No.</b> |
| Endotoxin Challenge Vial    | $\geq 2500$ EU/vial      | ECV2500V                 | 26055158       |
| <b>TEST (Method)</b>        | <b>SPECIFICATIONS</b>    | <b>Results</b>           |                |
| Endotoxin Potency (EU/vial) | $\geq 2500$ EU/vial      | Pass                     |                |
| <b>Storage</b>              | 2-8°C                    |                          |                |

Reconstitute ECV with 1.0 ml Water for BET, mix vigorously for at least 15 minutes on a vortex mixer.

Operator *Sandy xiao*

Date *2016. 05. 20*

QA Administrator *Jane Lin*

Date *2016. 05. 20*