



Endosafe[®]
Endotoxin Indicators (EVV10M)
Certificate of Quality

Each vial of Endotoxin Indicator Lot #: EVV94151 Expiration Date: Jan 2021

contains approximately 1,000,000 ng of lipopolysaccharide (derived from *E.coli* Strain 055:B5).

The biological activity is approximately 10,000,000 Endotoxin Units (EU) of lipopolysaccharide, as determined by serial two-fold dilution at the endpoint with Gel Clot U.S. Reference Lysate Lot # 15.

The absence of assay interfering substances was demonstrated by acceptable spike recovery in the range of 50% to 200%.

Qualified Analyst Christina Gusten Date 18 Jan 2019

Reviewed By Donna Sief Date 21 Jan 2019

Charles River Laboratories, Inc.
1023 Wappoo Road, Suite 43-B
Charleston, SC 29407 USA